





# **EU Type Examination Certificate**

This is to certify that: Prenger Healthcare B.V.

PO BOX 74 2280 AB Rijswijk The Netherlands

**Holds Certificate Number:** CE 751967

In respect of:

**Protective Half Masks** To EN 149:2001 + A1:2009 **See Continuation Sheets for Details** 

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Drs. Dave Hagenaars, Managing Director

First Issued: 2021-07-23 Latest Issue: 2021-07-23 Effective Date: 2021-07-23 Expiry Date: 2026-07-23

Page: 1 of 3



...making excellence a habit.™

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

## **EU Type Examination Certificate**

No. CE 751967

### **Product Specification**

**Product Type:** Filtering half masks to protect against particles.

Model(s): 2003

2004

2005 (mask with exhalation valve)

**Product description:** The particulate respirators are designed to protect against solid and non-volatile

> liquid particles. The masks are held on the face by a pair of elasticated straps. The models are single shift devices denoted by the classification symbol NR.

The Dolomite Clogging option is denoted by the classification symbol D.

**Technical specifications:** EN 149:2001+A1:2009 – Respiratory Protective Devices. Filtering half masks to

protect against particles.

Classification Model

**EN 149 Classification** 2003 FFP2 NR

> 2004 FFP2 NR D

> 2005 FFP3 NR D

First Issued: 2021-07-23 Effective Date: 2021-07-23 Latest Issue: 2021-07-23 Expiry Date: 2026-07-23

Page: 2 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

# **EU Type Examination Certificate**

No. CE 751967

#### **Certificate Administration Details**

Product initially approved on BSI certificates: CE 732331 and CE 739472

#### **Certificate Amendment Record:**

Issue date	Comments	<b>BSI Review Number</b>
July 2021	First issue.	2797:21:3476698

### **Certificate validity**

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall processes utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The validity of the Certificate for the products is also dependent on the maintenance of the EU Conformity to Type based on Internal Production Control plus supervised product checks at random intervals, Annex VII (Module D) as referenced on BSI issued Certificate CE 739473.

First Issued: 2021-07-23 Effective Date: 2021-07-23 Latest Issue: 2021-07-23 Expiry Date: 2026-07-23

Page: 3 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.

To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.